



# NABL

## SCOPE OF ACCREDITATION

Laboratory	Weightronics Mass Calibration Laboratory (WMCL), D-46, Sector-4, DSIIDC Industrial Area, Bawana, New Delhi		
Accreditation Standard	ISO/IEC 17025:2005		
Discipline	Mechanical Calibration	Issue Date	16.06.2016
Certificate Number	C-1403	Valid Until	15.06.2018
Last Amended on	23.11.2016	Page	2 of 2

Quantity Measured / Instrument	Range/ Frequency	* Calibration Measurement Capability ( $\pm$ )	Remarks
<b>II. WEIGHING SCALE AND BALANCE</b>			
<b>1. WEIGHING BALANCE<sup>S</sup></b>			
Readability: 10g	100 kg to 200 kg	20 g	Using Standard Weights of M1 Class (100 kg to 1000 kg) as per OIML R-76
Readability: 50g	500 kg to 1000 kg	100 g	

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95%

<sup>S</sup>Only in Permanent Laboratory

*S. Kunwar*

Sangeeta Kunwar  
Convenor

*Avijit Das*

Avijit Das  
Program Manager